Supplement

Manual Title:	1630-2/1630-2 FC Safety Information		
Part Number:	4811386	Supplement Issue:	**2**
Print Date:	October 2016	Issue Date:	5/20
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2017-2020 Fluke Corporation. All rights reserved.

FLUKE®

Change #1, 581

Under ***Safety Specifications***, change:

From:

Temperature Coefficient.....0.1 x (specified accuracy)/ °C (<18 °C or >28 °C)

To:

Temperature Coefficient.....0.15 % x reading/ °C (<18 °C or >28 °C)

From:

Dimensions.....283 mm x 107 mm x 48 mm

To:

Dimensions.....283 mm x 105 mm x 48 mm

Change #2, 67

In the ***Symbols*** table, remove:

Certified by TÜV SÜD Product Service.