Manual Supplement

Manual Title:	i410/i1010 Service Information	Supplement Issue:	**5**
Part Number:	642733	Issue Date:	4/10
Print Date:	December 1996	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

© 1999-2010 Fluke Corporation. All rights reserved. Printed in the U.S.A.

FLUKE®

Change #1

On page 3, under ***Required Equipment*** add the following:

> To ensure Range and Accuracy Specifications are maintained, warm up the Current Clamp for 20 minutes prior to performing the Performance Tests or Calibration and Adjustment procedures.

Change #2

On page 5, Table 4, replace steps 18 and 19 with the following:

18	18.0A dc	900.0A dc	0.9000V	NA	NA	881.5mV	918.5mV
19	-18.0A dc	-900.0A dc	-0.9000V	NA	NA	-918.5mV	-881.5mV

Change #3

On page 5, Table 4, step 23, remove 400Hz from the Calibrator Setting column.

Change #4

On page 5, Table 4, remove Step 21.

Change #5

On page 5, Table 4, remove Steps 8, 10 and 12.