Teledyne FLIR LLC
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: ***MR60***
Description: ***Moisture Meter Pro***
Date of Issue: ***12-Jul-21***

We, Teledyne FLIR LLC, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking

EMC Directive: **2014/30/EU**
Report Number: **11186580**
Report Date of Issue: **3/25/2016**

Standards:
EN 61326-1:2013

LVD Directive: **2014/35/EU**
Report Number: **E201687-A14-CB-1**
Report Date of Issue: **3/31/2016**

Standards:
IEC 61010-1 (ed.3)

RoHS Directive: **EU DIrective 2015/863/EU (RoHS 3)**

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive and EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive and EC EMC Directive and RoHS Directive.

Mark Sultzbach / QA Manager