# Manual Supplement

| | | | |
|---|---|---|---|
| **Manual Title:** | TL1000-MC4 Safety | **Supplement Issue:** | 1 |
| **Print Date:** | August 2022 | **Issue Date:** | 1/23 |
| **Revision/Date:** | | **Page Count:** | 1 |

This supplement contains information necessary to ensure the accuracy of the above document.

©2023 Fluke Corporation. All rights reserved.

FLUKE®

# Change #1

On page 2, replace the Symbols table with the following:

| Symbol | Description |
|---|---|
|  | Consult user documentation. |
|  | WARNING. RISK OF DANGER. |
|  | WARNING. HAZARDOUS VOLTAGE. Risk of electric shock. |
|  | Double insulated |
|  | Conforms to European Union directives. |
|  Intertek 5025476 | Conforms to UL STD 61010-031<br>Certified to CSA STD C22.2#61010-031 |
| **CAT III** | Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation. |
|  | This product complies with the WEEE Directive and its marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste. Do not dispose of this product as unsorted municipal waste. For information about take-back and recycling programs available in your country, see the Fluke website. |