Supplement

Title:	705 Inst.Sht.	Supplement Issue:	**4**
Print Date:	August 1998	Issue Date:	5/25
Revision/Date:	2, 6/03	Page Count:	2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	705 & 707
CD Rev. & Date:	1, 2/2008
CD PN:	2088974

© 2007-2025 Fluke Corporation.
All rights reserved. Printed in the U.S.A.

FLUKE®

Change #1

Under ***Specifications,*** change the ***Input/Output Protection***, to:

Input/output protection on mA range: self resetting, nonreplaceable 0.1 A fuse.

Change #2, 41512

Under ***General Specifications***, following the **Weight** specification, add:

Protection Class: Pollution Degree II

Change #3

On the second panel, under **International Symbols**, add the following:

N10140	Conforms to relative Australian standards.

Change #4, 57548

Under ***General Specifications***, replace the **Size** with:

Size: 69.85 mm W x 142.87 mm L x 38.10 mm H [2.75 in W x 5.625 in L x 1.50 in H]

With holster and Flex-Stand: 76.20 mm W x 158.75 mm L x 44.45 mm H [3.00 in W x 6.25 in L x 1.75 in H]

Change #5, TPUBS-335

Under ***Specifications***, ***DC Volts Input***, replace the Accuracy with:

±(0.0025 % + 2)