

Manual Supplement

Manual Title:	9142, 9143, 9144 Technical Guide	Supplement Issue:	2
Print Date:	January 2013	Issue Date:	3/18
Part Number:	Web-Only	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 488

On page 15, in the *Specification* table, replace the **TC Millivolt Range** with:

TC Millivolt Range	-10 mV to 100 mV
--------------------	------------------

Change #2, 539

On page 14, in the *Base Unit Specifications* table, replace the **Computer Interface** with:

Computer Interface	RS-232
--------------------	--------